

132 438



UNITED STATES
ENVIRONMENTAL PROTECTION AGENCY
REGION 5
CHICAGO, ILLINOIS

NOV 08 1990

DATE:

Review of Region 5 data for Schierco Service Co.
to, Director Chuck E. Hys
Region 5 Central Regional Laboratory
Data User:

Attached are the results for:

CRL Data Set Numbers: SF-7452
Sample Numbers: 91FF27548549, 247 SSQ, 91FF03257
Parameter(s): LOA / WATER
Laboratory: CRL / WESTON ESAT

Results Status:

- ☒ DATA ACCEPTABLE FOR USE
☒ DATA QUALIFIED AS TO USE for Vinyl chloride & vinyl acetate in
☐ DATA UNACCEPTABLE FOR USE all samples CMG 11/9/90

* For data acceptability requirements, refer to the method capability statement for the methods referenced.

Comments by the Quality Control Coordinator:

If there are any questions regarding the data, refer them to Steve Parker, James H. Adams,
the Quality Control Coordinator, at 353-3805.

Please sign and date this form below and return it with any comments to:

Sylvia Griffin
Data Management Coordinator
Region 5 Central Regional Laboratory
(SSCRL)

TRANSMITTED BY

NOV 08 1990

U.S. EPA CENTRAL
REGIONAL LAB

Cathy Kowis 11/13/90

RECEIVED BY/DATE:



UNITED STATES
ENVIRONMENTAL PROTECTION AGENCY
REGION 5
CHICAGO, ILLINOIS

DATE: NOV 08 1990
SUBJECT: Review of Region 5 data for Sobigro Service Co.
FROM: Curtis Ross, Director Chack & Eby
Region 5 Central Regional Laboratory
To: Data User:

Attached are the results for:

CRL Data Set Numbers: SF7452
Sample Numbers: 91FF07548, 549, D49, 550, 91FF03057
Parameter(s): ABN- Water
Laboratory: SCRL

Results Status:

- ☒ DATA ACCEPTABLE FOR USE*
☐ DATA QUALIFIED AS TO USE
☐ DATA UNACCEPTABLE FOR USE

* For data acceptability requirements, refer to the method capability statement for the methods referenced.

Comments by the Quality Control Coordinator:

Bis(2-ethylhexyl)phthalate was present in high concentration in all samples. This may indicate that it was introduced during extraction process. This compound is a common laboratory contaminant
C M G

If there are any questions regarding the data, refer them to ~~Steve Palmer~~, James H. Adams, the Quality Control Coordinator, at 353-3885.

Please sign and date this form below and return it with any comments to:

Sylvia Griffin
Data Management Coordinator
Region 5 Central Regional Laboratory
(SSCRL)

U.S. EPA CENTRAL
REGIONAL LAB

RECEIVED BY/DATE:

Comments:

Cathy Karris 11/15/90